

National Textiles & Garment Quality Supervision Inspection Centre (Zhejiang) Test Report

ADDRESS: No.300 XIASHA ROAD
HANGZHOUCHINA
Tel: (0571) 85122669
Fax: (0571) 85125121
Postcode: 310018

NUMBER: YJ202004990E
DATE: 2020-03-25

APPLICANT: ZHEJIANG YONGDA GARMENTS CO.,LTD
SAMPLE DESCRIPTION: Disposable Protective Masks
ADDRESS: MOJIALOU VILLAGE, TAOYAN TOWN, YUECHENG DISTRICT, SHAOXING CITY, ZHEJIANG PROVINCE CITY, ZHEJIANG PROVINCE

ITEM	STANDARD (D GRADE)	RESULT	RATING
1. FILTER EFFICIENCY(SALT MEDIUM) (GB/T32610-2016)(%)			
-	≥ 90	90.7	PASS
2. PARTICLE PROTECTIVE PERFORMANCE (GB/T32610-2016)(%)			
-	≥ 65	65.1	PASS
3. RESPIRATORY RESISTANCE (GB/T32610-2016)(Pa)			
-INSPIRATORY RESISTANCE	≤ 175	55	PASS
-EXPIRATORY RESISTANCE	≤ 145	46	PASS
4. THE BREAKING FORCE OF THE MASK BELT AND THE CONNECTION BETWEEN THE MASK BELT AND THE MASK BODY (GB/T 13773.2-2008)(GB/T32610-2016)(N)			
-	≥ 20	25	PASS

Remarks: 1. The protection effect grade is not marked on the sample mark, and it is judged according to the lowest standard requirement D grade.
2. The sample is not pretreated.

俞杰

PREPARED AND CHECKED BY
YU JIE

----END----